

## What is the probe measuring?

OxyProbes measure the partial pressure of oxygen. The partial pressure of Oxygen is dependent on the mixture of the gases and the pressure inside the vessel. Uncompensated variations in pressure can be a significant source of error. When reading % saturation or ppm, the operating pressure of the vessel should be entered into the transmitter if the process pressure is not the same as the calibration pressure. This number should account for elevation above sea level and/or any overpressure on the vessel. The pressure can be entered in mm Hg, inches Hg or bar.

---

The contents of this publication are presented for information purposes only, and while every effort has been made to ensure their accuracy, they are not to be construed as warranties or guarantees, express or implied, regarding products or services described herein or in their use or applicability. All sales are governed by our terms and conditions, which are available on request. We reserve the right to modify or improve the design or specification of such products at any time.

© 2022 Broadley-James Corporation. All rights reserved. Visit [www.broadley-james.com/trademarks](http://www.broadley-james.com/trademarks) for trademark information.

TMP-BF-102101



### North America and Pacific

Email: [sales@broadleyjames.com](mailto:sales@broadleyjames.com)  
Web: [www.broadleyjames.com](http://www.broadleyjames.com)

19 Thomas, Irvine CA, 92618 USA

Phone: 949-829-5555  
Tollfree: 800-288-2833  
Fax: 949-829-5560

### United Kingdom and EU Countries

Email: [sales@broadleyjames.co.uk](mailto:sales@broadleyjames.co.uk) Phone: +44 (0)1525 862518  
Web: [www.broadleyjames.eu](http://www.broadleyjames.eu) Fax: +44 (0)1525 862811

Wrest Park, Silsoe Beds MK45 4HS, UK