

Why Is The Gel In The Sensor Turning Brown?

Over time, the high temperatures of sterilization will discolor the gel in a FermProbe. This is often referred to as caramelisation. It is normal, and in no way affects the performance of the sensor. This color change can also be used to approximate the age of the sensor.

The contents of this publication are presented for information purposes only, and while every effort has been made to ensure their accuracy, they are not to be construed as warranties or guarantees, express or implied, regarding products or services described herein or in their use or applicability. All sales are governed by our terms and conditions, which are available on request. We reserve the right to modify or improve the design or specification of such products at any time.

© 2026 Broadley-James Corporation. All rights reserved. Visit www.broadley-james.com/trademarks for trademark information.

FAQ-225
TMP-BF-102101